

Guide to Using an UltraPak® without a Feeding Pump



1. Attach Bolink™ Small Cap to transfer formula from UltraPak® to a gravity bag.



2. If gravity bags are not available:

- Following clean technique, attach Bolink™ Small Cap to transfer formula safely from UltraPak® to Bolee™ Bag.
- Transfer up to 8 hours of formula (formula hang time in Bolee™ Bag is 8 hours).
- Connect Bolee™ Bag to the VESCO™ screw cap gravity set. Use drip chamber and roller clamp to control flow.
- Insert Bolee™ Bag into clear sleeve (included with VESCO™ screw cap gravity set) and hang from IV pole.
- Refill Bolee™ Bag every 8 hours until UltraPak® bag is empty (discard any formula remaining in UltraPak® bag after 48 hours).

Please note this information is NOT a substitute for clinical judgment regarding appropriate tube feeding administration.

For ordering information and inventory status, please contact the manufacturers: www.udelivermedical.com for Bolink™ Small Cap and Bolee™ Bag
www.vescomedical.com for VESCO™ MEDICAL Screw Cap Gravity Set

Bolink and Bolee are trademarks of U Deliver Medical, LLC. www.udelivermedical.com

Screw Cap gravity set is a product of VESCO Medical TM www.vescomedical.com

Except as noted, all trademarks are owned by Société des Produits Nestlé S.A., Vevey, Switzerland.

©2020 Nestlé. All rights reserved. Bridgewater, NJ 08807 U.S.A. 0420